



# VIDYASAGAR COLLEGE

NAAC ACCREDITED

39, SANKAR GHOSH LANE, KOLKATA-700 006

PHONE : (Principal) - (033) 2241 3018, (Office) - (033) 2241 4447

e-mail : vidyasagarcollegekolkata@gmail.com ★ Website : www.vidyasagarcollege.edu.in

Ref. No. VC/Tender/2024/11

Date 11.09.2024

Tender Ref No.....

Dated.....

Tender for Quotation valid from 11.09.2024 to 18.09.2024

## OPEN TENDER NOTICE

Sealed Quotations are invited from companies having trade licence, PAN Card, GST and Professional Tax declaration (Submit Xerox copies for same) for purchase/work of Magnetic stirrer (1 unit) and Digital pH Meter (1 unit) as per the given specification for the Department of Zoology (PG section) under the aegis of Department of Science & Technology & Biotechnology, Government of West Bengal, Biotechnology based Opportunities Offered to Science & Technology Departments BOOST Programme 2017-18, 1097/BT(Estt)/1P-08/2018, Dated Dated:24.10.19)

Specifications/details of the items required are mentioned below:

- Price must be quoted including GST
- The quoted price must be in rounded form
- The top of the envelope for quotation must specify Name of the Department, Tender Ref No. and name of the College as Vidyasagar College.
- Incomplete quotations will be rejected without any clarification. Discretion of Tender & Purchase Committee, Vidyasagar College would be final in case of any dispute.

Quotations are to be placed in the Drop-box in front of the College Office within the valid date

Sl No.	Item	Specifications	Quantity/any other
1	Magnetic stirrer	With Hot Plate, Speed control (max 1800 rpm)	1 (one)
2	Digital pH Meter	High degree of accuracy, 0.01pH, Error indication for buffers & slope, Auto/Man. temp. compensation, High input impedance, Auto polarity & decimal, Compact & light weight, Displays pH/mV/temp., pH Mode: -- Range: 0.00 to 14.00 pH, - Resolution: 0.01 pH, -Accuracy: 0.01 pH $\pm$ 1 digit, mV Mode: - - Range: 0 to $\pm$ 1000 mV, -Resolution: $\pm$ 1 mV -Accuracy: 1 mV $\pm$ 1 digit, - Input Impedance: $10^{12}$ ohms, - Buffer calibration: 7.00 pH, 4.00 pH and 9.20 pH, -Temp. range: 0-150°C with continuous digital display, -Resolution : 1.0°C, -Temp. compensation : Automatic & Manual, liquid crystal display.	1 (one)

*[Signature]* 10/09/24

Coordinator, BOOST  
Dept. Of Zoology

*[Signature]* 10/09/24

Convenor, T/P Committee  
Vidyasagar College  
Tender & Purc. Committee  
Vidyasagar College  
Kolkata-700 006

*[Signature]* 10/9/24

Principal  
Vidyasagar College  
Principal  
Vidyasagar College  
Kolkata-700006